



Certificate of Analysis

Customer Information

Client: Alkaloid Consumers
Attention: admin@alkaloidconsumers.com
Address: 5753 S Prince St, #7
Littleton, CO 80120

Testing Facility

Lab: Cora Science, LLC
Address 8000 Anderson Square, STE 113
Austin, Texas 78757
Contact: info@corascience.com
(512) 856-5007

Sample Image(s)



Sample Information

Name: Bronx Extracts 88.1% MGM
Lot Number: NA
Description: Powdered fine chemical
Condition: Good
Job ID: ISO06546
Sample ID: I18306
Received: 13MAR2026
Completed: 16MAR2026
Issued: 17MAR2026

Test Results

Assay (w/w%) [UHPLC-DAD]		Method Code: T100		Tested: 14MAR2026 0300		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Dihydro-7-hydroxymitragynine	Report Results	91.6	w/w%	0.085	N/A	

Residual Solvents: Class I (GC-MS)		Method Code: T201		Tested: 16MAR2026 1641		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
1,1-Dichloroethene	NMT 8	<LOQ	ug/g	0.40	PASS	
1,1,1-Trichloroethane	NMT 1500	<LOQ	ug/g	75	PASS	
Tetrachloromethane	NMT 4	<LOQ	ug/g	0.20	PASS	
Benzene	NMT 2	<LOQ	ug/g	0.10	PASS	
1,2-Dichloroethane	NMT 5	<LOQ	ug/g	0.25	PASS	

Residual Solvents: Class II (GC-MS)		Method Code: T201		Tested: 16MAR2026 1641		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Methanol	NMT 3000	208	ug/g	75	PASS	
Acetonitrile	NMT 410	<LOQ	ug/g	41	PASS	
Dichloromethane	NMT 600	<LOQ	ug/g	15	PASS	
1,2-Dichloroethene, (E)	NMT 1870	<LOQ	ug/g	46.75	PASS	
1,2-Dichloroethene, (Z)	NMT 1870	<LOQ	ug/g	46.75	PASS	
Tetrahydrofuran	NMT 720	<LOQ	ug/g	18	PASS	
Cyclohexane	NMT 3880	<LOQ	ug/g	97	PASS	
Methylcyclohexane	NMT 1180	<LOQ	ug/g	29.5	PASS	
1,4-Dioxane	NMT 380	<LOQ	ug/g	38	PASS	
Toluene	NMT 890	<LOQ	ug/g	22.25	PASS	
Chlorobenzene	NMT 360	<LOQ	ug/g	9	PASS	

This report, prepared by Cora Science, LLC, shall not be reproduced except in its entirety without prior written approval. All test articles are analyzed as received and the results relate only to the specific sample of material or product analyzed. Test methods are performed in a laboratory accredited to ISO/IEC 17025:2017 in the field of testing by PJLA (Accreditation #116374) or a registered outsourcing facility. Some test methods reported may fall outside the scope of L22-250 supplement.

PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES
Ethylbenzene	NMT 2170	<LOQ	ug/g	54.25	PASS
o/p-Xylene	NMT 2170	<LOQ	ug/g	54.25	PASS
m-Xylene	NMT 2170	<LOQ	ug/g	54.25	PASS
Isopropylbenzene	NMT 70	<LOQ	ug/g	1.75	PASS
Hexane	NMT 290	<LOQ	ug/g	7.25	PASS
Nitromethane	NMT 50	<LOQ	ug/g	1.25	PASS
Chloroform	NMT 60	<LOQ	ug/g	1.5	PASS
1,2-Dimethoxyethane	NMT 100	<LOQ	ug/g	2.5	PASS
Trichloroethene	NMT 80	<LOQ	ug/g	2	PASS
Pyridine	NMT 200	<LOQ	ug/g	5	PASS
2-Hexanone	NMT 50	<LOQ	ug/g	5	PASS
Tetralin	NMT 100	<LOQ	ug/g	2.5	PASS
Total Xylenes	NMT 2170	<LOQ	ug/g	54	PASS

Residual Solvents: Class III (GC-MS)		Method Code: T201		Tested: 16MAR2026 1641		
PARAMETER	SPECIFICATION	RESULT	UNIT	LOQ	NOTES	
Pentane	NMT 5000	<LOQ	ug/g	125	PASS	
Ethanol	NMT 5000	<LOQ	ug/g	125	PASS	
Diethyl Ether	NMT 5000	<LOQ	ug/g	125	PASS	
Acetone	NMT 5000	<LOQ	ug/g	125	PASS	
Ethyl Formate	NMT 5000	<LOQ	ug/g	125	PASS	
Isopropanol	NMT 5000	<LOQ	ug/g	125	PASS	
Methyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS	
Methyl tert-Butyl Ether	NMT 5000	<LOQ	ug/g	125	PASS	
1-Propanol	NMT 5000	<LOQ	ug/g	125	PASS	
2-Butanone	NMT 5000	<LOQ	ug/g	125	PASS	
Ethyl Acetate	NMT 5000	874	ug/g	125	PASS	
2-Butanol	NMT 5000	<LOQ	ug/g	125	PASS	
2-Methyl-1-Propanol	NMT 5000	<LOQ	ug/g	125	PASS	
Isopropyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS	
Heptane	NMT 5000	<LOQ	ug/g	125	PASS	
1-Butanol	NMT 5000	<LOQ	ug/g	125	PASS	
Propyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS	
4-Methyl-2-Pentanone	NMT 5000	<LOQ	ug/g	125	PASS	
Isoamyl Alcohol	NMT 5000	<LOQ	ug/g	125	PASS	
Isobutyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS	
1-Pentanol	NMT 5000	<LOQ	ug/g	125	PASS	
Butyl Acetate	NMT 5000	<LOQ	ug/g	125	PASS	
Dimethylsulfoxide	NMT 5000	<LOQ	ug/g	125	PASS	
Anisole	NMT 5000	<LOQ	ug/g	125	PASS	

Additional Report Notes

N/A

Revision History

Report ID: 851d0edf-64a1-4e3c-a2e2-d026e6fee7ef
rev 00 - Initial release.

Abbreviations

ID: identification, **N/A:** not applicable, **LOQ:** limit of quantitation, **CFU:** colony forming units, **w/w%:** weight by weight percent, **mg:** milligrams, **g:** grams, **ug:** micrograms, **mL:** milliliters, **ND:** not detected, **<LOQ:** below limit of quantitation, **NMT:** no more than, **NLT:** no less than, **UHPLC:** ultra-high performance liquid chromatography, **GC:** gas chromatography, **DAD:** diode array detection/detector, **MS:** mass spectroscopy/spectrometer, **ICP:** inductively coupled plasma, **ISO:** International Organization for Standardization, **USP:** United States Pharmacopeia

Authorization

This report has been authorized for release from Cora Science by:

Signature:		Position:	Laboratory Director
Name:	Tyler West	Department:	Management
		Date:	17MAR2026